





**Certificate:** 80199450

**Master Contract:** 304611

**Project:** 80199450

**Date Issued:** 2024-08-19

**Conditions of Acceptability:**

1. External measuring probes are not part of the certification.
2. Connection and disconnection to/from mains supply shall be evaluated in the end use application.
3. Overcurrent protection shall be evaluated in the end use application.
4. The equipment shall be supplied by a power source 10 to 24 Vdc with double or reinforced isolation from mains which is secondary limited (LPS or Class2).
5. The equipment is only to be installed in a separate cabinet/enclosure (built-in) to protect against mechanical influences.

**APPLICABLE REQUIREMENTS**

CSA C22.2 No. 61010-1-12, UPD1:2015, UPD2:2016 AMD1:2018 - Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 1: General requirements...

CSA C22.2 No. 61010-2-030:18 - Safety requirements for electrical equipment for measurement, control, and laboratory use — Part 2-030: Particular requirements for equipment having testing or measuring circuits - Second Edition

UL 61010-1 3rd ed (Rev. Nov 21, 2018) - UL Standard for Safety Electrical Equipment For Measurement, Control, and Laboratory Use; Part 1: General Requirements - Third Edition; Including Revisions through November 21, 2018

UL 61010-2-030, 2nd Edition (2018) - UL Standard for Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use – Part 2-030: Particular Requirements for Equipment Having Testing or Measuring Circuits



**Certificate:** 80199450


**Master Contract:** 304611

**Project:** 80199450

**Date Issued:** 2024-08-19

Notes:

Products certified under Class(es) C363106 have been certified under CSA's ISO/IEC 17065 accreditation with the Standards Council of Canada (SCC). [www.scc.ca](http://www.scc.ca)





## *Supplement to Certificate of Compliance*

**Certificate:** 80199450

**Master Contract:** 304611

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
80199450	2024-08-19	Initial cCSAus Component Certification of Cleverscope CS548 HV Oscilloscopes according to 61010-1 3rd Ed. + AMD1 based on CB Acceptance (provided by Dekra) and partial re-testing by CSA Group.